

UNISONICS

DATA AND SPECIFICATIONS

ULTRASONIC CLEANER MODEL ST18M

APPLICATIONS:

The Unisonic ST range is a single chamber device used to clean surgical instruments and other hardware with sonic energy in a heated water/detergent solution. It is designed for use in surgery reprocessing areas, central processing departments and laboratories.

FEATURES:

- Solid stainless steel construction (tank & jacket).
- Mechanical timer.
- 600 watts of ultrasonic power.
- Sweep frequency.
- Polyester coated transducers.
- Stainless steel lid and basket (standard in all models).
- Insulation.
- Outlet drain.
- Cooling fan.
- Overflow apron.
- EMC tested and approved.

SONIC CLEANING CHAMBER

Free Standing Cabinet

Overall Size	(External)	Length 470mm Width 470mm Height 500mm
Chamber Size	(Internal)	Length 300mm Width 300mm Depth 200mm
Basket Size		Length 250mm Width 250mm Depth 160mm
Maximum Load		10kg
Chamber Volume		18 litres
Operating liquid capacity		14 litres

APPROVALS:

AS/NZS3760:2001

AS/NZS2064:1997-EMI Test

AS3100-1994-Electrical Safety

Certificate of Conformity No:E990002-(C-Tick)

Certificate for inclusion of medical device – Class 1 (T.G.A.)

UNISONICS

DATA AND SPECIFICATIONS

ULTRASONIC CLEANER MODEL ST18M

CONSTRUCTION:

All exterior surfaces are 0.9mm in 304 stainless steel.

The ultrasonic chamber is fabricated from 304 stainless steel and is 1.6mm in thickness.

It is heat insulated and sound deadened.

All integral plumbing is stainless steel.

Basket is stainless steel mesh.

Sonic cleaning energy is provided to the chamber by three piezoelectric transducers bonded to the tank bottom with frequency of 40 kHz.

Maximum operating temperature should not exceed 60 degrees Celsius so as to maintain reliability and maximise efficiency.

Power is supplied by 600 watts of solid-state circuitry, which is air-cooled.

CONTROLS:

A mechanical timer control is located at the bottom front of the unit and should be set as per operating instructions (refer operation page).

A circuit breaker is also fitted for circuit protection and personal safety.

The mains plug/socket outlet is located at the side of the unit, which allows the operator to remove the connection when the cleaner is not in use.

A fan is located on the opposite side to allow cross cooling and ventilation of power circuitry.

The unit should always be positioned to allow for plenty of circulation.

A valve-draining outlet is located at the front of the unit and care should always be taken to not allow excessive spillage of solution when draining fluid.

GENERAL CLEANING:

The turnover so solution must be determined by the user to satisfy acceptable cleaning results of the finished article.

The more the solution is contaminated the longer the cleaning process.

If the contamination is heavy and difficult to remove a pre-rinse in a heated bath could be sufficient enough to soften and loosen unwanted debris in preparation for the ultrasonic cleaning.

At the other end of the scale after the items have been removed it is essential that a final wash be performed to remove any residue that remains on the cleaned parts.